

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

ST- Win Basic is a Windows based Stress Test System. It enables accurate computation of ST measurements, complying full frequency response of 0.05 Hz to 100 Hz

Technical Specifications

1. Environmental

Operating Temp : -5 °C to +50 °C

Operating & Storage : 0 to 95 % non-condensing

Humidity

2. Power supply

Input : 230V AC+/- 10% @ 50Hz

Power consumption : < 10 Watt

3. Performance

Display : 1024 x 768 pixels,

Language : English

Alarm : Audio alarms for HR and ST level at pre-configured
Values

QRS Beep : On each QRS

4. Acquisition

Frequency response : 0.05 Hz to 100 Hz with notch at 50Hz


Filter : DSP

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

Input impedance	: 10 MΩ
Time constant	: > 3.2 sec
A/D conversion	: 12 bit
Noise	: < 20μV _{p-p}
Leakage current	: < 10 μA
Patient isolation	: Optical isolation.
Gain	: 5, 10 & 20 mm/mV selectable.
CMRR	: > 120 dB
Sweep speed	: 25.0 mm/S.
Accuracy of HR	: +/- 2 BPM.
Sampling Frequency	: 250 samples per second simultaneous
Model	: STWin-Standard, STWin-Basic, STWin-Deluxe
Storage	: 12 lead unaveraged ECG, 10 second Rhythm strips of 12 leads, Medians after every 8 sec along with HR, BP and fiducial points, full disclosure of entire test
Protocols	: Pre-programmed protocol including BRUCE, MODIFIED BRUCE, NAUGHTON, MODIFIED BALKE and user editable protocols
Printouts	: Automatic and manual online printing with grid On/off facility, Print colour reports

5. Treadmill


Power	: 230V, 50Hz, 15A
-------	-------------------

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

Power consumption	: 1.5KVA
Incline Range	: 0-22%
Dimension	: L X W - 2080 X 785 mm
Walking Area	: 1350 X 500mm
Weight	: 100Kg
Patient Weight	: Up to 250kg
Belt Speed	: Smooth, 0-15Kmph
In-House manufacturing	: Yes
Independent earthing connection	
Servo Stabilizer 5KV for treadmill	

6. Minimum Requirement of PC for ST-Win system:

1. Operating system : Windows 10 professional 64bit
2. Motherboard : Intel corei3
3. RAM : 4 GB
4. USB port : working USB port for Simul-G unit connection.
5. Monitor : Color monitor.
6. Hard disk : minimum 500 GB.
7. Display Settings : Resolution - 1024X 768.
8. Keyboard : with 101 keys.
9. Mouse
10. Independent earthing connection.
11. CD ROM : Make : LG. Samsung Sony, ASUS.
12. Printer : HP Laser Jet 1000, HP Laser Jet 1010
- 13.1 OR 2 KVA Online UPS for PC and treadmill controller

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

Detailed features

❖ Acquisition

- Body level acquisition of 12 lead ECG
- Simultaneous acquisition of 12 Leads of ECG @ 250 samples per second per channel with ADC resolution of 12-bit.
- Frequency Response – DSP filter of 0.05 to 100Hz with notch at 50Hz
- Facility to store 12 lead unaveraged ECG
- Facility to record rhythm (10 second Rhythm strips of 12 leads)
- Beep on QRS
- Facility to Rerun test

❖ Acquisition display modes

- **3 lead display**
 - 8 seconds display of 2 leads of ECG + 1 QRS detection Lead.
 - Facility to configure 3 leads to any of the 12 ECG leads
 - Display lead sets (I, II, III) OR (aVR, aVL, aVF) OR (V1, V2, V3) OR (V4, V5, V6)
 - An enlarged Exercise median of the minimum ST level lead with fiducial points marked is displayed. (The enlarged median can be any of the 12 leads in Manual Mode. In Automatic mode, the lead having maximum ST Depression is displayed by the system.)
 - Along with enlarged median, basal median is displayed in exercise stage and peak exercise median is displayed in recovery for comparison
- **Online display of following parameter in all display modes**
 - METS
 - Heart rate is updated every 4 seconds with symbol and Target heart rate with percentage of completion target heart rate.
 - BP in mmHg
 - ST measurement Mode (Manual / Auto)
 - Current ST Level and ST Slope of enlarged median
 - ECG gain with standardization pulse
 - Test and Phase duration in hh:mm format
 - Exercise protocol stage name

❖ Online printouts

- Facility to print selected online reports
 - Linked Median report (12 linked medians with ST Levels of selected long lead)
 - 12 lead Raw ECG report

❖ Acquisition settings

- Facility to enter BP 50 seconds before end of each exercise stage
- Relearn the median template
- Facility to enter a stage comment in each phase
- Gain: 0.5, 1.0, 2.0 mV/cm

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

- Change the QRS detection lead
- Automatic/Manual detection of fiducial points (E, J, Post J)
- Hold/Release treadmill stage
- QRS beep On/Off
- NASAN filters ON/OFF

❖ **Review**

- Review data of selected patient
- Grid/Graph/No grid option on all review screens
- Facility to edit HR and BP of any stage
- Facility to edit Patient Details

❖ **Review display modes**

- 12 medians screen
 - 12 medians are displayed with ST-level
- Mixed Median report with one Median complex + 2.5 seconds ECG of each lead
- Facility to zoom median of selected lead and edit fiducial points
- Trends
 - Trends of HR, BP, ST Level, ST Slope, J Amplitude
 - 8, 16, 32, 48, 56 seconds resolution.
- View and print rhythm strips marked during acquisition. Facility to select 12 leads of 5 sec. OR 6 leads of 10 seconds for each printout
- View unaveraged ECG
- Edit and print Summary Report using standard edit options

❖ **Reports**

- 12 lead linked median (single / all stages)
- Mixed Median Report
- 6 lead Rhythm report of 10 seconds (single / all stages)
- 12 lead Unaveraged ECG report
- Extra Comments report
- Stage Report
- Brief Summary Report
- Trends Report (HR, BP, ST Levels / ST Slopes, J Amplitude of 3 configured leads)
- Summary Report (Hospital Address and referring doctor designation to be printed in Summary reports)
- ST-Levels & Slopes Table
- Selected median report

❖ **Settings**

- Hospital details
- Referring Doctor list
- Standard Summary report option
- ST-Win system configuration
- Treadmill protocol
 - Available standard protocols
 - BRUCE
 - MODIFIED BRUCE
 - NAUGHTON
 - MODIFIED BALKE
- Printing details configuration

	Title: Website Catalog & specification Document	Document no: NASAN-D&D-D-007
		Rev no: 5
		Date: 29-1-2024
		Details of change: new format

- System configuration
 - Speed in km/hr or m/s
 - TM speed at exercise stop (slow-0 km/hr, Fast-0 km/hr / 1.2 km/hr)
 - ST Level lead in Summary
 - Long lead in reports
 - Screen Display leads (3)
 - 3 Leads in Trend Report
 - Stage Report Leads
 - Treadmill Support: Compact, Treadstar, TreadstarXP, Trackmaster
 - COM Port Selection for amplifier.
 - Median update time is configurable to 1 to 8 sec.

❖ Utilities

- Test Treadmill
- Test acquisition unit
- Calibrate TM Speed
- Calibrate TM Grade
- Delete printout from printer queue

❖ Display

- Resolution – 1024 X 768
- color monitor

*Due to our continuous product development program, the specification may change without prior notice.